

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 11/25/2013

mmB

NSL

Instrument Serial Number: 010568

As of 25-Nov-13

Start Date	End Date	Agency Name
22-Nov-13		DFS Central Lab
29-Apr-13	22-Nov-13	Prince William County PD
12-Mar-13	29-Apr-13	DFS Central Lab
21-Nov-11	12-Mar-13	Richmond PD
17-Nov-11	21-Nov-11	DFS Central Lab
07-Sep-11	17-Nov-11	Richmond PD
12-Feb-10	07-Sep-11	DFS Central Lab
16-Apr-09	12-Feb-10	Virginia Beach PD
23-Jun-08	16-Apr-09	DFS Central Lab

mmB  
WHL

## INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010568 Worksheet Start Date 11/22/2013

Location Prince William Garfield PD

Address 15948 Donald Curtis Dr Woodbridge VA 22191

DFS Technician Michelle Billeter License No. 26417

☐ Laboratory ☒ On-Site

### Site Specifications

	Yes	No
<b>1 Secured Environment:</b> Location must not be accessible to the general public.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
<b>2 Environment free from contaminants</b>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
<b>3 Dedicated telephone line:</b> The phone line should be used solely by the instrument.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
<b>4 Adequate Ventilation:</b> Enough space must be provided for the instrument and the printer to allow for adequate ventilation.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
<b>5 Stable temperature controls:</b> The instrument should not be exposed to direct sources of temperature variation.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
<b>6 115VAC power source:</b> Whenever possible instrument power circuit should not be shared with refrigerators, copy machines or other high amperage devices.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
<b>7 Room Condition:</b> The site should be free of excessive moisture, heat and dust.	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Old serial no.	New serial no.	Supplies
E-Seek		Mouthpieces
Keyboard		Certificates of Analysis
Printer		Operator Worksheet
UPS		Other:

Maintenance (if any)

Comments (if any)

COULD NOT PERFORM INSTRUMENT TEST OR MEASUREMENT ASSURANCE DUE TO INSTRUMENT BEING DISABLED. REMOVED TO DFS CENTRAL.

mmB  
11/26/13

Instrument Serial Number 010568

Certification Date

Instrument Barometer (mm HG) 763

Reference Barometer (mm HG)

765

Reference Barometer(RB)Serial # 009111

RB Calibration Due

5/23/2014

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

all measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

☐ Calibrated☐ Certified☐ Replaced dry gas standard (+O-ring)

Dry gas standard Lot No. (with tank no.)

AG320502-11

☐ Measurement Assurance Check☐ Run a Breath Test (Instrument Test)

(successful completion denotes satisfactory condition of the item)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

DFS Technician

*Michelle Bell*

Date

11/25/13

Issuing Analyst

*Mehmet Kenedy*

Date

11/26/13

Intox EC/IR II QA Worksheet

Issued by Calibration and Training Program Manager

Issue Date: 4-March-2013

TLN

DFS Document 250-F113

Revision 6

Page 2 of 2